



- ① Multi-purpose clip
- ② Holder for positioning stick in duct
- ③ 0.5 oz Dia.  
The humidity sensor is protected by a quick twist of the bottom of the probe stem → the protective cap is closed. Should only be left open for the duration of the measurement.
- ④ Swivel head. The display can always be read.

2 sec. before pressing the button again. The selected unit then applies.

### Technical data

Type of application:	.....Short measurement
Measuring range:	.....5 to 95 %RH / +32 to +122 °F / -4 to +122 °F td
Resolution:	.....0.1 %RH / 0.1 °F
Accuracy:	.....(System accuracy at 25 °C / +77 °F) ±3 %RH / ±0.9 °F
Parameters:	.....%RH, °C, °F
Ambient temperature:	.....32 to +122 °F
Storage temperature:	.....-4 to +160 °F
Battery type:	.....3 x AAA
Battery lifetime:	.....Approx. 1000 h
Probe stem:	.....0.5 in
Length:	.....4.9 in
Warranty:	.....2 years
CE guideline	.....2004/108/EWG

### Commissioning

Remove the protective film on the display.

### Switching on

- On Press button:



Segmenttest  
testo 605-H1



Segmenttest  
testo 605-H2




Current reading

### Changing the parameters

- On Press button repeatedly:



Temperature



Dew point  
testo 605-H1



Wet bulb  
testo 605-H2



Humidity

### Switching off

- On Keep button pressed for 3 seconds.

### Selecting between °C/°F

- On When switching on keep button pressed until the segment test is finished, instrument switches to configuration menu. Press button again the instrument alternates between °C and °F. Wait

### Please read before using instrument

- Avoid condensation
- Observe measurement ranges of sensor
- Do not exceed maximum storage and operating temperatures (e.g. protect measuring instrument from direct sunlight).
- Inexpert handling cancels your warranty.

### Changing the battery



The battery has to be changed if the symbol appears in the display during the measurement.

- ✓ Instrument is switched off.
  - 1 Open battery compartment cover.
  - 2 Remove spent batteries and insert new batteries (3x AAA) into the battery compartment. Observe polarity!
  - 3 Close battery compartment cover.

### Auto Off function

- ⏸ If no button has been pressed in approx. 10 minutes, the instrument switches itself off automatically.